

C E R T I F I C A T E O F C O M P L I A N C E

LOT #: 644205

KEY SCIENTIFIC PRODUCTS
1113 E REYNOLDS STREET
STAMFORD, TEXAS 79553
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: PYR DISCS (W/PEP REAGENT)
CATALOG #: K1538B

LOT #: 644205

PACKAGED DATE = 2024/08/20 EXP DATE = 2026/08/20

CONTROL LOT# 5761

TESTED BY: BLC

REAGENT USED-PEP:

		CONTROL	THIS LOT	
13813	STREPTOCOCCUS AGALAC		NEG	NEG
29212	ENTEROCOCCUS FAECALI		POS	POS
19615	STREPTOCOCCUS PYOGEN		POS	POS
25923	STAPHYLOCOCCUS AUREU		NEG	NEG
13882	KLEBSIELLA PNEUMONIA		POS	POS
13047	ENTEROBACTER CLOACAE		POS	POS
25922	ESCHERICHIA COLI		NEG	NEG
13048	ENTEROBACTER AEROGEN		POS	POS
8100	SERRATIA MARCESSENS		POS	POS
12386	STREPTOCOCCUS, GROUP		NEG	NEG

CONTROL STRIPS 1-16
ALL STRIPS PASSED

POSITIVE TEST IS RED OR DARK PINK AFTER ADDING
REAGENT. GREEN, YELLOW, OR VERY PALE PINK IS NEGATIVE.
TESTED CONTROL STRIP FROM EACH TUB AND RANDOM DISCS.