

C E R T I F I C A T E O F C O M P L I A N C E

LOT #: 635502

KEY SCIENTIFIC PRODUCTS
1113 E REYNOLDS STREET
STAMFORD, TEXAS 79553
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: PYR DISCS (W/PEP REAGENT)
CATALOG #: K1538B

LOT #: 635502

PACKAGED DATE = 2024/01/05 EXP DATE = 2026/01/05

CONTROL LOT# 630500

TESTED BY: BLC

REAGENT USED-PEP: 632400 , 624800, 600500

	CONTROL	THIS LOT	
13813	STREPTOCOCCUS AGALAC	NEG	NEG
29212	ENTEROCOCCUS FAECALI	POS	POS
19615	STREPTOCOCCUS PYOGEN	POS	POS
25923	STAPHYLOCOCCUS AUREU	NEG	NEG
13882	KLEBSIELLA PNEUMONIA	POS	POS
25922	ESCHERICHIA COLI	NEG	NEG
13048	ENTEROBACTER AEROGEN	POS	POS
8100	SERRATIA MARCESSENS	POS	POS
12386	STREPTOCOCCUS, GROUP	NEG	NEG

CONTROL STRIPS 1-26 -PASSED
ALL STRIPS PASSED

POSITIVE TEST IS RED OR DARK PINK AFTER ADDING
REAGENT. GREEN, YELLOW, OR VERY PALE PINK IS NEGATIVE.
TESTED CONTROL STRIP FROM EACH TUB AND RANDOM DISCS.