

C E R T I F I C A T E   O F   C O M P L I A N C E

LOT #: 648303

KEY SCIENTIFIC PRODUCTS  
1113 E REYNOLDS STREET  
STAMFORD, TEXAS 79553  
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED  
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED  
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: BUTYRATE DISCS  
CATALOG #: K9131B

LOT #: 648303

PACKAGED DATE = 2025/02/20    EXP DATE = 2027/02/20

-----  
CONTROL LOT# 636400

TESTED BY: BLC  
-----

		CURRENT POSITIVE	CURRENT NEGATIVE	CONTROL POSITIVE	CONTROL NEGATIVE
13090	NEISSERIA MENINGITID		COLORLESS		COLORLESS
19424	NEISSERIA GONORRHOEA		COLORLESS		COLORLESS
25923	STAPHYLOCOCCUS AUREU	BLUE		BLUE	
25240	BRANHAMELLA CATARRHA	BLUE		BLUE	
49143	MORAXELLA CATARRHALI	BLUE		BLUE	
19615	STREPTOCOCCUS PYOGEN		COLORLESS		COLORLESS
23970	NEISSERIA LACTAMICA		COLORLESS		COLORLESS

DRY PAPER COLOR - WHITE  
POSITIVE IS BLUE AFTER INCUBATION OF 15 MINUTES  
NEGATIVE REMAINS COLORLESS